

510(k) Summary
S-Scan
Esaote S.p.A.

16100620

APR - 5 2010

510(k) Summary

The following 510(k) summary has been prepared pursuant to requirements specified in 21CFR 807.92(a).

Submitter Information

Donatella Ragghianti
Esaote, S.p.A.
Via Siffredi 58
Genova 16153 Italy

Contact Person: Allison Scott, RAC
P: 317.569.9500 x106
F: 317.569.9520
ascott@ansongroup.com

Date: March 5, 2010

Trade Name: S-Scan

Common Name: System, Nuclear Magnetic Resonance Imaging

Classification Name(s): Magnetic Resonance Diagnostic Device

Classification Number: 90LNH

Predicate Device(s)

Tradename	Common name	Class	Product code	Manufacturer	K number
S-scan	System, nuclear magnetic resonance imaging	II	LNH	ESAOTE S.P.A.	K080968

Device Description

S-scan is a Magnetic Resonance (MR) system, which produces images of the internal structures of the patient's limbs and joints.

This 510(k) is to add the DPA cervical coil. This is a dual phased array receiving coil, shaped for suitability to the areas under examination and designed to be matched to the patient's cervical vertebrae.

Intended Use(s)

S-scan is a Magnetic Resonance (MR) system that produces transverse, sagittal and coronal and oblique cross-section images of the limbs, joints and spinal column. It is intended for imaging portions of the upper limb, including the hand, wrist, forearm, elbow, arm and shoulder, imaging portions of the lower limb, including the foot, ankle, calf, knee, thigh and hip and imaging portions of the spinal column, including the cervical and lumbo-sacral sections.

S-scan images correspond to the spatial distribution of protons (hydrogen nuclei) that determine magnetic resonance properties and are dependent on the MR parameters, including spin-lattice relaxation time (T1), spin-spin relaxation time (T2), nuclei density, flow velocity and "chemical shift". When interpreted by a medical expert trained in the use of MR equipment, the images can provide diagnostically useful information.

Technological Characteristics

The addition of the DPA Cervical Spine Coil to the S-scan system, reflected in this Special 510(k), does not alter the fundamental scientific technology of the S-scan system, the predicate device, cleared via K080968.



DEPARTMENT OF HEALTH & HUMAN SERVICES

Public Health Service

Food and Drug Administration
10903 New Hampshire Avenue
Document Control Room – WO66-G609
Silver Spring, MD 20993-0002

APR - 5 2010

Esaote S.p.A.
% Ms. Allison Scott, RAC
Regulatory Associate
Anson Group, LLC
11460 N. Meridian Street, Suite 150
CARMEL IN 46032

Re: K100670

Trade/Device Name: Esaote S-Scan MR System
Regulation Number: 21 CFR 892.1000
Regulation Name: Magnetic resonance diagnostic device
Regulatory Class: II
Product Code: MOS
Dated: March 5, 2010
Received: March 9, 2010

Dear Ms. Scott:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.

If your device is classified (see above) into class II (Special Controls), it may be subject to such additional controls. Existing major regulations affecting your device can be found in Title 21, Code of Federal Regulations (CFR), Parts 800 to 895. In addition, FDA may publish further announcements concerning your device in the Federal Register.


Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Parts 801 and 809); medical device reporting (reporting of

medical device-related adverse events) (21 CFR 803); and good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820). This letter will allow you to begin marketing your device as described in your Section 510(k) premarket notification. The FDA finding of substantial equivalence of your device to a legally marketed predicate device results in a classification for your device and thus, permits your device to proceed to the market.

If you desire specific advice for your device on our labeling regulation (21 CFR Parts 801 and 809), please contact the Office of *In Vitro* Diagnostic Device Evaluation and Safety at (301) 796-5450. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address <http://www.fda.gov/cdrh/industry/support/index.html>.

Sincerely yours,

A handwritten signature in black ink, appearing to read "Donald J. St. Pierre", written over a horizontal line.

Donald J. St. Pierre
Acting Director
Division of Radiological Devices
Office of *In Vitro* Diagnostic Device
Evaluation and Safety
Center for Devices and Radiological Health

Enclosure

Indications for Use

510(k) Number (if known): **k100670**

Device Name: Esaote S-Scan MR System

Indications for Use:

"S-scan is a Magnetic Resonance (MR) system that produces transversal, sagittal and coronal and oblique cross-section images of the limbs, joints and spinal column. It is intended for imaging portions of the upper limb, including the hand, wrist, forearm, elbow, arm and shoulder, imaging portions of the lower limb, including the foot, ankle, calf, knee, thigh and hip and imaging portions of the spinal column, including the cervical and lumbo-sacral sections.

S-scan images correspond to the spatial distribution of protons (hydrogen nuclei) that determine magnetic resonance properties and are dependent on the MR parameters, including spin-lattice relaxation time (T₁), spin-spin relaxation time (T₂), nuclei density, flow velocity and "chemical shift". When interpreted by a medical expert trained in the use of MR equipment, the images can provide diagnostically useful information."


Prescription Use X
(Part 21 CFR 801 Subpart D)

AND/OR

Over-The-Counter Use _____
(21 CFR 807 Subpart C)

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of In Vitro Diagnostic Devices (OIVD)



(Division Sign-Off)

Division of Radiological Devices

Office of *In Vitro* Diagnostic Device Evaluation and Safety

510(k) Number k100670